



# National Agri-Food Biotechnology Institute

Department of Biotechnology, Government of India

C-127, Industrial Area, S.A.S. Nagar,

Mohali-160071 Punjab, INDIA

Ph: 0172-2290300

Fax: 0172-4604888

NABI/ 1 (111)/ 11-12/ N-PUR

23<sup>rd</sup> Jan' 2012

## NOTICE INVITING QUOTATION

Sealed Quotations in TWO BID SYSTEM are invited on behalf of Executive Director NABI, Mohali for the purchase of following items so as to reach latest by 12<sup>th</sup> Feb' 2012 on or before 3.30 pm. Technical and Commercial bids should be sent in separate sealed envelopes and enclosing them in a common sealed cover. The Quotations will be opened on the same day at 4 pm in the presence of tenderers, if any:

### Instrument name: Gloss Meter

**Qty – 1 No**

### Technical specifications

- 1) A device is required for simultaneous measurement of gloss at three different angles, 20, 60 & 85°.
- 2) Measurement range: 0-1800 GU (20°); 0-800 GU (60°); 0-150 GU (85°)
- 3) Mode of measurement: actual value, average, minimum, maximum
- 4) Certified calibration standard tile should be the standard scope of supply.
- 5) Memory: at least 900 reading
- 6) Interface and software: USB and any other advanced interface
- 7) Language: English
- 8) All accessories required for successful installation, functioning and regular maintenance of the equipment must be quoted.
- 9) Warranty: 2 years
- 10) Multiple models with higher specifications should be quoted as optional.
- 11) The instrument must meet international standards of quality certification.
- 12) Bidders should give a compliance statement (point wise) showing/highlighting items part no/serial number in the quotation, covering letter, or attached technical brochure, for comprehensive technical comparison. Proof of compliance should be mentioned point wise in the catalogue. Failing in compliance and proof of compliance will cause cancellation of the bid without any further notice/information. **"Please fill the format as per Annexure A"**
- 13) Multiple Models with higher specifications should be quoted as separate models in the bids.

BEFORE QUOTING PLEASE REFER THE INSTRUCTIONS OVERLEAF CAREFULLY

Yours faithfully,

-sd-

(S. Krishnan)

Stores and Purchase Officer